



C.E DECLARATION OF CONFORMITY

COMPANY: BHM MEDICAL INC.
 ADDRESS: 2001 Tanguay
 Magog, Qc, Canada
 J1X 5Y5

TELEPHONE: 819-868-0441
 FAX: 819-868-2249

HEREBY DECLARE THAT:

PRODUCT: FLOOR LIFT
 MODEL: MINISTAND

ARE ACCORDING TO THE STANDARDS LISTED BELOW AMONG OTHERS:

<u>NUMBER</u>	<u>TITLE</u>	<u>DATE</u>
EN 60601-1-1	Electromedical Apparatus - Security Standards	2002
EN 60601-1-2	Electromedical Apparatus - Security Standards and Electromagnetic Compatibility	2002
EN 1041	Terminology and Symbols	2002
EN 1441	Medical Appartus - Risk Analysis	2002
EN 12182	Technical Aids for Disabled Persons - General Requirements and Tests Methods	2002
ISO 10535:1998	Hoist for the Transfert of Disabled Person - Requirement and Tests Methods	2002

AND IS ACCORDING TO THE FOLLOWING RULES AND GUIDE LINES:

<u>NUMBER</u>	<u>TITLE</u>
93/42/CEE	Medical Rules and Guide Lines
93/68/CEE	C.E. Marking Guide Lines

SIGNATURE: 

DATE: 1/28/2003

NAME: BERNARD J. HAMEL, c.a.
 VICE-PRESIDENT

MADE IN:
 MAGOG, QUEBEC, CANADA